Middlesex University Research Repository

An open access repository of
Middlesex University research

<http://eprints.mdx.ac.uk>

de Gelidi, Serena ORCID logo ORCID: <https://orcid.org/0000-0001-6141-2736>, Seifnaraghi, Nima, Bardill, Andy ORCID logo ORCID: <https://orcid.org/0000-0002-2451-3260>, Wu, Yu, Frerichs, Inéz, Demosthenous, Andreas, Tizzard, Andrew ORCID logo ORCID: <https://orcid.org/0000-0002-6159-4901> and Bayford, Richard ORCID logo ORCID: <https://orcid.org/0000-0001-8863-6385> (2020) Towards a thoracic conductive phantom for EIT. Medical Engineering and Physics, 77 . pp. 88-94. ISSN 1350-4533 [Article] (doi:10.1016/j.medengphy.2019.10.008)

Final accepted version (with author's formatting)

This version is available at: <https://eprints.mdx.ac.uk/28763/>

Copyright:

Middlesex University Research Repository makes the University's research available electronically.

Copyright and moral rights to this work are retained by the author and/or other copyright owners unless otherwise stated. The work is supplied on the understanding that any use for commercial gain is strictly forbidden. A copy may be downloaded for personal, non-commercial, research or study without prior permission and without charge.

Works, including theses and research projects, may not be reproduced in any format or medium, or extensive quotations taken from them, or their content changed in any way, without first obtaining permission in writing from the copyright holder(s). They may not be sold or exploited commercially in any format or medium without the prior written permission of the copyright holder(s).

Full bibliographic details must be given when referring to, or quoting from full items including the author's name, the title of the work, publication details where relevant (place, publisher, date), pagination, and for theses or dissertations the awarding institution, the degree type awarded, and the date of the award.

If you believe that any material held in the repository infringes copyright law, please contact the Repository Team at Middlesex University via the following email address:

eprints@mdx.ac.uk

The item will be removed from the repository while any claim is being investigated.

See also repository copyright: re-use policy: <http://eprints.mdx.ac.uk/policies.html#copy>

Towards a thoracic conductive phantom for EIT

Serena de Gelidi^{a,*}, Nima Seifnaraghi^a, Andy Bardill^a, Yu Wu^b, Inéz Frerichs^c,
Andreas Demosthenous^b, Andrew Tizzard^a, Richard Bayford^a

^a*Faculty of Science & Technology, Middlesex University, London, United Kingdom*

^b*University College London, London, United Kingdom*

^c*University Medical Centre Schleswig-Holstein, Kiel, Germany*

Number of words: 4854

Number of words in the abstract: 128

---

*Corresponding author

Email address: s.degelidi@mdx.ac.uk (Serena de Gelidi)

Abstract

Phantom experiments are a crucial step for testing new hardware or imaging algorithms for electrical impedance tomography (EIT) studies. However, constructing an accurate phantom for EIT research remains critical; some studies have attempted to model the skull and breasts, and even fewer, as yet, have considered the thorax. In this study, a critical comparison between the electrical properties (impedance) of three materials is undertaken: a polyurethane foam, a silicone mixture and a thermoplastic polyurethane filament. The latter was identified as the most promising material and adopted for the development of a flexible neonatal torso. The validation is performed by the EIT image reconstruction of the air filled cavities, which mimic the lung regions. The methodology is reproducible for the creation of any phantom that requires a slight flexibility.

Keywords: phantom, EIT, 3D printing, conductive, flexible

1. Introduction

Tissue-Mimicking Materials (TMM) are required to test and validate diverse emerging biomedical applications [1], including Electrical Impedance Tomography (EIT). EIT seeks to reconstruct the changes in impedance distribution within tissues caused by related physiological activities. This is achieved by acquiring data from injecting a set of currents into the body through surface electrodes and measuring the boundary voltages [2].

Definition

Phantoms are objects meant to replicate the main features of the final application of the device and associated anatomy under consideration. Phantom experiments are the bridge between computer-based simulations and clinical measurements, as they investigate the performance of the developed data-acquisition system, reconstruction algorithms or imaging software and subsequently provide reasonable information for further optimization [3]. Furthermore, phantoms allow a controlled *in vitro* testing, which is difficult to achieve in clinical experiments. A review of materials selected in previous works to generate a phantom is detailed in the following sections and summarised in Table 1.

The most common option

Saline-filled tanks, made of insulating materials and featuring metallic electrodes, are usually adopted to perform pilot measurements for EIT reconstructions [4, 5, 6]. Isaacson et al. [7] introduced anatomical features by suspending agar cross-sections of heart and lungs in a saline circular tank.

Flexible tanks and phantoms

Aiming to acquire boundary deformation measurements, Boyle et al. [8] designed a sponge rubber ring featuring stainless steel electrodes. As a result, the phantom was easy to compress yet its elasticity ensured the return to its original shape. Belmont et al. [9] showed that cylindrical tofu specimens were a

28 viable phantom for soft TMM compression studies utilizing bioimpedance tech-
 29 niques. The main limitation of adopting food organic materials as TMM is
 30 their perishable nature. Despite the fact that several anthropomorphic resusci-
 31 tation manikins are available to clinicians for training purposes, none of these
 32 can mimic the skin conductivity as they are made of insulating polymers. Since
 33 EIT can be adopted for different purposes, previous works have attempted the
 34 generation of an improved *in vitro* setup compared to the circular tanks. Tiz-
 35 zard et al. [10] used a gelatine breast phantom to ultimately generate a more
 36 accurate forward model. However, agar and gelatines degrade over time in con-
 37 tact with air or water, making them unsuitable for sporadic use over the long
 38 term [11].

39 *Anatomically realistic phantoms*

40 Recently, the idea of the common cylindrical tank has been upgraded into a
 41 geometrically accurate skull [11]. Hence, Avery et al. [11] 3D printed the skull
 42 model by means of a polylactic acid (PLA) filament and filled it with saline
 43 solution. As the adopted PLA is non conductive, they also needed to place
 44 33 electrodes against the tank walls. Dunne et al. [12] employed a conduc-
 45 tive material for generating the first anatomically accurate pelvic phantom for
 46 EIT. The chosen TMM was obtained by mixing the composition of 30% w/w
 47 graphite powder, 5.7% w/w carbon black (CB) powder and the remainder from
 48 equal parts of polyurethane precursors [1, 13]. A similar recipe was adopted to
 49 fabricate a two-layers head phantom for use in EIT [14] and breast tumours for
 50 Microwave Imaging [15]. Zhang et al. [3] created, based on 3D printing tech-
 51 niques, a novel four-layers structure head phantom with anatomically realistic
 52 geometry and continuously varying skull resistivity. Two types of acrylonitrile
 53 butadiene styrene (ABS) and CB particles with volume fractions of 10% and
 54 20% CB were fabricated [3]. Similarly, Kurrant et al. [16] 3D printed a variety
 55 of sizes and shapes intended as breast models to test a prototype estimating
 56 the surface for Microwave Imaging. However, they did not specify the material
 57 used. Burfeindt et al. [17] prepared a 3D printed breast phantom made of ABS

for use in microwave breast-imaging experiments. A different approach was attempted by Garrett and Fear [13], who created 3D printed molds to pour in a hand-made mixture of CB, graphite and rubber mixture [1]. Faenger et al. [18] were among the first to propose the application of conductive 3D printable filaments. Therefore, they claimed that conductive ABS or PLA are sensible choices. Their breast phantom for Microwave Imaging consisted in two interior 3D printed containers and a silicone composite based skin, which was created by mixing silicone, CB and graphite [18].

66

[Table 1 about here.]

Aim of the study

Therefore, the construction of an accurate phantom for EIT research remains a critical and challenging step. As part of the CRADL project (<http://cradlproject.org/>), which is developing EIT technology as supportive method for monitoring neonatal ventilation, the need of a phantom to test the prototypes arose. More generally, the phantom could help to improve the development of EIT and to establish it as a bedside method for optimising ventilation therapy [19]. The present study aims to develop a conductive and flexible neonatal phantom to test prototypes for the thoracic boundary detection [20] and EIT reconstruction. In order to generate a phantom, three materials have been selected and analysed in terms of both their electrical impedance properties and their production technique.

2. Materials and methods

2.1. Materials

The main requirements selected for developing the thoracic prototypes are the electrical conductivity and the elasticity, meaning that the mechanical behaviour of each material should feature low stiffness. Hence, three different materials have been selected to be compared:

A) A carbon impregnated polyurethane foam (Teknis Limited, UK).

- 86 B) A mix of a silicone (75%), CB powder (15%) and graphite powder (10%)
 87 [18], which needs to be synthesized. CB powder has been preferred over
 88 carbon fibres in order to promote the isotropy of the generated material.
- 89 C) A carbon filled thermoplastic polyurethane Palmiga 95-250 (Creative Tools,
 90 Sweden) has been acquired among the newest conductive and flexible fil-
 91 aments available for 3D printing.

92 2.2. Preparation

93 Five samples of each material were prepared, featuring the same cross-section
 94 (10mm x 10mm) and the following lengths: 10 mm, 20 mm, 30 mm, 40 mm and
 95 50 mm. Different sample lengths were tested for linearity and homogeneity of
 96 the material, to record the variation in impedance. Successively, phantoms of
 97 idealized geometry were prepared in order to mimic the dimensions of a neonatal
 98 torso featuring a diameter of 7.5 cm and two lungs, simplified as through holes
 99 (Figure 1).

100
 101 Material *A* is available in a ready to use form. However, since the foam is
 102 manufactured and sold in sheet form, it was cut by an abrasive water jet cutter.

103
 104 Using similar methodology to Garrett and Fear [13], material *B* was prepared
 105 by hand in a fume hood due to the toxicity of CB in powder form. Hence, the
 106 rubber solution (prior to curing) was weighed and mixed in a container by hand.
 107 The CB powder was then weighed, added to the rubber mixture, and the mate-
 108 rials were mixed with a metal stirrer for several minutes [1]. The mixture was
 109 prepared in accordance with the percentages reported by Garrett and Fear [1] to
 110 mimic the skin conductivity: 63 wt% silicone, 7 wt% CB, 30 wt% graphite. This
 111 choice is justified by the need to mimic a neonatal torso, in which the bone and
 112 the fat properties were assumed not to be overall predominant. However, the
 113 curing of the material *B* proved to be unsuccessful when using both a platinum-
 114 cure silicone Transil 20 (Mouldlife, UK) and a water white clear urethane Clear
 115 Flex 30 (Smooth-On, US). The same result was observed even when changing

the percentages to the ones reported by Faenger et al. [18] for skin TMM: 75 wt% silicone, 15 wt% CB, 10 wt% graphite. However, such percentages were successfully adopted in the preparation of material *B* by the use of tin cure silicone TinSil Gel-10 (Polytek, US). Hence, powders were weighed and mixed. The two parts of the rubber solution were also weighed and mixed in a container by hand for a couple of minutes. Powders were added to the compound, which was stirred. In order to remove the majority of air bubbles the mixture was subjected to ultrasonication and then immediately poured in the custom-made mould.

Lastly, samples of the material *C* were 3D printed by means of a Printrbot Simple Metal (Printbot, US). In order to obtain flexible samples, the infill was kept as low as 20%. Preliminary tests showed that printing below such value of infill led to suboptimal results. However, given the overall limited quality obtained and the fact that the other printers available were not compatible with such filament, the 3D print of the phantom was commissioned externally, keeping the same infill percentage.

[Figure 1 about here.]

2.3. Testing

Samples of each material were tested by means of a Solartron 1260 impedance analyzer (Solartron Analytical, UK) in order to compare their electrical properties. The absolute permittivity ϵ_{abs} has been measured by sweeping the frequency up to 2 MHz, which is the band of interest for biological tissues in the EIT field [21, 14]. The relative permittivity ϵ_r of each sample was calculated as: $\epsilon_r = \epsilon_{abs}/\epsilon_0$, where the vacuum permittivity ϵ_0 is approximated to $8.85 \cdot 10^{-12} \text{ F} \cdot \text{m}^{-1}$.

Following the characterization of the sample material, the contact impedance of each phantom was tested by means of an EIT setup. The system, shown in

Figure 2, mimics the EIT belt applied along the chest circumference of neonates as part of the CRADL project. The belt prototype was made of 32 copper tape electrodes placed on a PVC substrate. Salt-free electrode gel (Spectra 360, Parker Laboratories, US) was applied on the interface between the belt electrodes and the phantom. The raw measurements were recorded by the Pioneer Set (SenTec AG, CH) and processed in a Matlab (The MathWorks, US) custom script. The reference was taken by filling the holes of the model with cylinders of analogous electrical properties. In order to mimic the air content in the lungs, the main measurement featured the holes empty.

[Figure 2 about here.]

2.4. Image reconstruction

The idealized phantom model has been meshed in COMSOL Multiphysics (COMSOL Inc, SE) by means of 10,594 tetrahedral elements and exported to Matlab. Successively, the inverse model has been created by means of the GREIT algorithm [22] using the EIDORS v 3.9 toolbox (<http://eidors3d.sourceforge.net/>). The image is reconstructed using a difference EIT method [23] with the reference measurement, during which the through holes of the idealized phantom were filled by insertions made of the same material. Therefore, the reference domain is homogeneous.

2.5. The neonatal phantom

A thoracic model of a term baby, gestational age 38 weeks, has been previously developed based on the CT scans and ultrasound images [24]. Such a CAD model has been selected to create an anatomically accurate phantom of torso. The section between the shoulder height and the diaphragm has been chosen to identify clearly the armpit, as the anatomical landmark below which the EIT belt is usually placed by nurses. Hence, the CAD model has been cut, for the sake of practicality, parallel to the transverse plane. In addition, since the CRADL project focus is on lung ventilation, the model has been simplified by including only the lungs among the inner parts. The concavities of the lungs

174 have been neglected in order to simplify the design. Lastly, the lungs have been
 175 designed as entities removable from the torso in order to mimic the air content
 176 by removing them prior to recording the reference data. Such a design forms
 177 the basis of further development of the phantom leading to more complex *in*
 178 *vitro* testing.

179 3. Results

180 3.1. Testing the samples

181 The average ϵ_r of the three materials, which was calculated after measuring
 182 the $\epsilon_{absolute}$ of each sample up to 2 MHz, is shown in Figure 3. The average ϵ_r
 183 of the 3D printed samples, made of material *C*, is always higher compared to
 184 the other materials. Materials *A* and *B* show the same permittivity above 100
 185 KHz, while below such threshold the silicone B exhibits lower resistance to the
 186 electric field (Figure 3). The effect of the methodology adopted to prepare the
 187 samples is clearly reflected in the electrical properties: the hand-prepared mix
 188 leads to a high variability in the standard deviation, indicated by the blue error
 189 bars in Figure 3.

190

191 [Figure 3 about here.]

192 3.2. Testing the phantoms

193 The phantoms were tested by means of the EIT setup, shown in Figure 2,
 194 firstly without the use of electrode gel. The average contact impedances of all
 195 three materials were around 1460 Ω , being far too high for the application. This
 196 is motivated by the fact that the Pioneer Set has an impedance limit for the
 197 current source, defined as 700 Ω maximum. The application of the electrode
 198 gel on the interface lowered the contact impedance to 1100 Ω for material *A*,
 199 1000 Ω for material *B* and 500 Ω for material *C*. Therefore, it was possible to
 200 carry out an EIT analysis only of the 3D printed phantom.

201

202 3.3. Image reconstruction

203 The EIT reconstruction of the 3D printed phantom was achieved by means
204 of the GREIT algorithm [22]. As shown in Figure 4, the cavities were detected
205 with the correct orientation in reference to the first electrode, which in EIDORS
206 is represented in a lighter green compared to the others.

207 [Figure 4 about here.]

208 Given the results obtained, the neonatal phantom (Section 2.5) has been 3D
209 printed by means of the material C , as shown in Figure 5, and tested similarly
210 to the idealized one (Figure 2). The average contact impedance was $1410\ \Omega$ in
211 absence of the electrode gel and lowered to $185\ \Omega$ when it was applied.

212 [Figure 5 about here.]

213 The measured voltages have been imported in Matlab in order to reconstruct
214 the EIT image. The 3D geometry has been meshed in Comsol resulting in a
215 model comprising 20,640 tetrahedral elements and 4483 nodes (Figure 6).

216 [Figure 6 about here.]

217 The blue areas shown in Figure 7 B highlight the detection of the lung
218 cavities in the corresponding orientation reported in Figure 7 A.

219 [Figure 7 about here.]

220 4. Discussion

221 The present study has compared, for the first time, three materials for the
222 preparation of a flexible and conductive phantom to be used for EIT *in vitro*
223 testing. The sample preparation has highlighted the strengths and limitations
224 of each material. The analysis below critically reviews each option. For ex-
225 ample, material A has low elasticity and is therefore difficult to shape to the
226 desired geometry by cutting. Moreover, as the material is only available in sheet

227 form, larger phantoms would require many sheets to be laminated, presenting
 228 discontinuities in the model. Material *B* was obtained after attempting the
 229 polymerization with several silicones. The curing inhibition, which surprisingly
 230 has not been experienced in any previously reported work, could be related to
 231 impurities (e.g. sulphur) contained in the graphite. Therefore, it is suggested
 232 that the recipe used by Garrett and Fear [13], McDermott et al. [14] and Dunne
 233 et al. [12] may be polymerized only by using the specific polyurethane VytaFlex
 234 20 (Smooth-On, US). In contrast to the foam, no limitation in terms of dimen-
 235 sions is associated with material *B* as long as a specific mould is prepared in
 236 advance. However, the manual preparation compromises the homogeneity of
 237 such a dense compound, which is reflected in the electrical properties (Figure
 238 3). Although the recipe of Faenger et al. [18] allowed the preparation of the
 239 silicone mix and it was successfully used for Microwave Imaging, its resulting
 240 ϵ_r was about the same as that of the air in the tested frequency range (Figure
 241 3). While the 3D printing process is extremely versatile, only selected printers
 242 can handle material *C*. Furthermore, the nozzle on the 3D printer can easily
 243 become clogged and adversely affected by the abrasive nature of carbon. Hence
 244 a hardened steel nozzle is commonly recommended.

245 Due to the contact impedance cut off of the Pioneer Set, materials *A* and *B* are
 246 deemed not viable for EIT purposes. This appears to be in contrast with Dunne
 247 et al. [12], however these authors reported a higher percentage of graphite for
 248 the pelvic phantom and they filled the cavity with sodium chloride and ultra-
 249 sound gel.

250 The recorded values of contact impedance were $500\ \Omega$ and $185\ \Omega$ when electrode
 251 gel was applied to the 3D printed idealized and neonatal phantoms respectively.
 252 These values are close to the average of $300\ \Omega$ observed by Sophocleous et al.
 253 [25] in preterm infants after applying the neonatal ultrasound gel to the EIT
 254 belt (SenTec AG, CH) interface. Similarly, the median value of skin contact
 255 impedance observed in adults after applying the ContactAgent (SenTec AG,
 256 CH) on the textile EIT SensorBelt (SenTec AG, CH) was $325\ \Omega$ [26]. However,
 257 a wider range of contact impedances was recorded in this study compared to

that undertaken by Sophocleous et al. [25]. The authors attribute such difference to the type of the electrodes, being textile in Sophocleous et al. [25] and made of simple copper tape in the present work.

Time-difference image reconstructions were successfully carried out using finite element models matching the experimental 3D printed phantom-prototype setup. Figures 4 and 7 show a change in conductivity in the correct location where the phantoms are air filled, respectively in the idealized and neonatal phantoms. Thus, it can be concluded that a carbon filled thermoplastic polyurethane phantom is a viable option for testing EIT prototypes. The additional advantages of such material are the possibility of generating a patient specific shape by using a 3D printer, as well as being flexible enough to emulate changes in lung shape over time.

Although the anatomical features for the neonatal torso and the human skull are different, this research confirms the accuracy in terms of location and shape of the target area detected as the electrical impedance imaging results are similar to the ones obtained by Zhang et al. [3], who generated and tested a 3D printed four-layer skull model made of ABS and CB. At present this article presents the first example of a neonate phantom for EIT. The use of a thermoplastic polyurethane increases the mechanical flexibility of the phantom, thus simulating the anatomical similarity. Unfortunately, mechanical properties of conductive filaments are rarely quantified by suppliers. According to the ISO 527, material *C* has an elongation at break of 250%, while carbon fibre ABS and PLA feature a value of 2% (CarbonX CF, 3DXTECH, US). Such ease of stretching of the selected material was modulated by the infill of the printing in order to mimic the torso flexibility.

Material homogeneity and the simplified internal design of the phantoms with the lungs being the sole internal components, represent the main limitations of the present work. In addition, even though the layer deposition of the printing process is consistent with the ribs' orientation and leads to an anisotropic behaviour of the phantom, this aspect is clearly simplified.

Overall, the material analysis carried out in this study will support the devel-

opment of phantoms for other applications of EIT. In particular, the flexibility is a critical aspect to be replicated in several anatomically realistic applications, where a deformable boundary would lead to a different EIT image. The neonatal thorax is more compliant compared to the adult one by anatomical composition [27]. Future work will be undertaken to increase the complexity of the neonatal phantom, as the experimental evaluation of a prototype system is a necessary precursor to its clinical use. The authors aim to model different test scenarios, including diseased regions of the lungs.

5. Conclusion

Among the selection of materials and production techniques explored, the carbon filled thermoplastic polyurethane was favourably validated for the fabrication of an anatomically correct, conductive and flexible phantom. The 3D printing ensures the material homogeneity and the customization of the internal structure. A simplified neonatal torso of a phantom was thus generated. Similarly to the clinical practice, raw voltages were collected by means of the Pioneer set after applying a layer of ultrasound gel to the phantom surface. The corresponding image reconstruction showed the correct location of the air filled cavities, which feature a negative change of conductivity given the above-mentioned material adopted for the phantom. The novel results obtained in this study are therefore highlighting the possibility to standardise the *in vitro* testing of the EIT device by using a known material before facing the huge biological variability of the clinical practice. Lastly, such phantom would be helpful to attempt different designs of belt without trying it on humans.

Acknowledgements

This work is supported by the CRADL project which has received funding from the European Unions Horizon 2020 research and innovation programme 2014-2018 under grant agreement No 668259.

316 Competing interests: None declared.

317 Ethical approval: Not required.

318 References

- 319 [1] J. Garrett, E. Fear, Stable and flexible materials to mimic the dielectric
320 properties of human soft tissues, *IEEE Antennas and Wireless Propagation*
321 *Letters* 13 (2014) 599–602. doi:10.1109/LAWP.2014.2312925.
- 322 [2] R. Bayford, Bioimpedance Tomography (Electrical Impedance Tomog-
323 raphy), *Annual Review of Biomedical Engineering* 8 (2006) 63–91.
324 URL: [http://www.annualreviews.org/doi/10.1146/annurev.bioeng.](http://www.annualreviews.org/doi/10.1146/annurev.bioeng.8.061505.095716)
325 [8.061505.095716](http://www.annualreviews.org/doi/10.1146/annurev.bioeng.8.061505.095716). doi:10.1146/annurev.bioeng.8.061505.095716.
- 326 [3] J. Zhang, B. Yang, H. Li, F. Fu, X. Shi, X. Dong, M. Dai, A novel 3D-
327 printed head phantom with anatomically realistic geometry and continu-
328 ously varying skull resistivity distribution for electrical impedance tomog-
329 raphy, *Scientific Reports* 7 (2017) 4608. URL: [http://www.nature.com/](http://www.nature.com/articles/s41598-017-05006-8)
330 [articles/s41598-017-05006-8](http://www.nature.com/articles/s41598-017-05006-8). doi:10.1038/s41598-017-05006-8.
- 331 [4] A. P. Bagshaw, A. D. Liston, R. H. Bayford, A. Tizzard, A. P. Gibson,
332 A. T. Tidswell, M. K. Sparkes, H. Dehghani, C. D. Binnie, D. S. Holder,
333 Electrical impedance tomography of human brain function using recon-
334 struction algorithms based on the finite element method, *NeuroImage* 20
335 (2003) 752–764. doi:10.1016/S1053-8119(03)00301-X.
- 336 [5] B. Grychtol, B. Müller, A. Adler, 3D EIT image reconstruction with
337 GREIT., *Physiological Measurement* 37 (2016) 785–800. URL: [http://](http://www.ncbi.nlm.nih.gov/pubmed/27203184)
338 www.ncbi.nlm.nih.gov/pubmed/27203184. doi:10.1088/0967-3334/37/
339 6/785.
- 340 [6] C. Busch, B. Schullcke, K. Moeller, Realization of a multi-layer EIT-system,
341 *Current Directions in Biomedical Engineering* 3 (2017) 291–294.

- [7] D. Isaacson, J. L. Mueller, J. C. Newell, S. Siltanen, Reconstructions of chest phantoms by the D-bar method for electrical impedance tomography, IEEE Transactions on Medical Imaging 23 (2004) 821–828. doi:10.1109/TMI.2004.827482.
- [8] A. Boyle, A. Adler, W. R. B. Lionheart, Shape Deformation in Two-Dimensional Electrical Impedance Tomography 31 (2012) 2185–2193.
- [9] B. Belmont, R. E. Dodde, A. J. Shih, Impedance of tissue-mimicking phantom material under compression, Journal of Electrical Bioimpedance 4 (2013) 2–12. URL: <https://www.journals.uio.no/index.php/bioimpedance/article/view/443>. doi:10.5617/jeb.443.
- [10] A. Tizzard, Borsic, R. Halter, R. H. Bayford, Generation and performance of patient-specific forward models for breast imaging with EIT, Journal of Physics: Conference Series 224 (2010) 012034. URL: <http://stacks.iop.org/1742-6596/224/i=1/a=012034?key=crossref.39d32895f1b56224f9f42bee4635446a>. doi:10.1088/1742-6596/224/1/012034.
- [11] J. Avery, K. Aristovich, B. Low, D. Holder, Reproducible 3D printed head tanks for electrical impedance tomography with realistic shape and conductivity distribution, Physiological Measurement 38 (2017) 1116–1131. doi:10.1088/1361-6579/aa6586.
- [12] E. Dunne, B. McGinley, M. O’Halloran, E. Porter, A realistic pelvic phantom for electrical impedance measurement, Physiological Measurement 39 (2018) 034001 (10pp). doi:<https://doi.org/10.1088/1361-6579/aaa3c0>.
- [13] J. Garrett, E. Fear, A New Breast Phantom With a Durable Skin Layer for Microwave Breast Imaging, IEEE Transactions on Antennas and Propagation 63 (2015) 1693–1700. doi:10.1109/TAP.2015.2393854.

- [14] B. McDermott, B. McGinley, K. Krukiewicz, B. Divilly, M. Jones, M. Biggs, M. O'Halloran, E. Porter, Stable tissue-mimicking materials and an anatomically realistic, adjustable head phantom for electrical impedance tomography, *Biomedical Physics & Engineering Express* 4 (2017) 015003. URL: <http://stacks.iop.org/2057-1976/4/i=1/a=015003?key=crossref.5a638d936a7418d4d3f2a8c9b86e583b>. doi:10.1088/2057-1976/aa922d.
- [15] B. L. Oliveira, D. O'Loughlin, M. O'Halloran, E. Porter, M. Glavin, E. Jones, Microwave Breast Imaging: experimental tumour phantoms for the evaluation of new breast cancer diagnosis systems, *Biomedical Physics & Engineering Express* 4 (2018) 025036. URL: <http://stacks.iop.org/2057-1976/4/i=2/a=025036?key=crossref.2405731769d3a91297cf3e51ebcb3394>. doi:10.1088/2057-1976/aaaaff.
- [16] D. Kurrant, J. Bourqui, E. Fear, Surface estimation for microwave imaging, *Sensors (Switzerland)* 17 (2017). doi:10.3390/s17071658.
- [17] M. J. Burfeindt, T. J. Colgan, R. O. Mays, J. D. Shea, N. Behdad, B. D. Van Veen, S. C. Hagness, MRI-Derived 3-D-Printed Breast Phantom for Microwave Breast Imaging Validation, *IEEE Antennas Wirel Propag Lett.* 11 (2012) 1610–1613. doi:10.1016/j.jmb.2008.10.054. The arXiv:NIHMS150003.
- [18] B. Faenger, S. Ley, M. Helbig, J. Sachs, I. Hilger, Breast phantom with a conductive skin layer and conductive 3D-printed anatomical structures for microwave imaging, in: *11th EUCAP*, volume 1, 2017, pp. 1065–1068. doi:10.23919/EuCAP.2017.7928662.
- [19] I. Frerichs, M. B. P. Amato, A. H. van Kaam, D. G. Tingay, Z. Zhao, B. Grychtol, M. Bodenstein, H. Gagnon, S. H. Böhm, E. Teschner, O. Stenqvist, T. Mauri, V. Torsani, L. Camporota, A. Schibler, G. K. Wolf, D. Gommers, S. Leonhardt, A. Adler, Chest electrical impedance tomography examination, data analysis, terminology, clinical use and recommenda-

- 398 tions: consensus statement of the TRanslational EIT developmeNt stuDy
399 group, *Thorax* 72 (2017) 83–93. doi:10.1136/thoraxjnl-2016-208357.
- 400 [20] S. de Gelidi, N. Seifnaraghi, A. Bardill, A. Tizzard, Y. Wu, E. So-
401 rantin, S. Nordebo, A. Demosthenous, R. Bayford, Torso shape
402 detection to improve lung monitoring, *Physiological Measurement*
403 39 (2018) 074001. URL: [http://stacks.iop.org/0967-3334/39/i=7/](http://stacks.iop.org/0967-3334/39/i=7/a=074001?key=crossref.57a6cbf0eccc142bed13a34efc47650d)
404 [a=074001?key=crossref.57a6cbf0eccc142bed13a34efc47650d](http://stacks.iop.org/0967-3334/39/i=7/a=074001?key=crossref.57a6cbf0eccc142bed13a34efc47650d). doi:10.
405 1088/1361-6579/aacc1c.
- 406 [21] B. Rigaud, J.-P. Morucci, N. Chauveau, Bioelectrical impedance techniques
407 in medicine Part I: Bioimpedance measurement second section: Impedance
408 spectrometry, *Critical Reviews in Biomedical Engineering* (1997).
- 409 [22] A. Adler, J. H. Arnold, R. Bayford, A. Borsic, B. Brown, P. Dixon, T. J.
410 Faes, I. Frerichs, H. Gagnon, Y. Gärber, B. Grychtol, G. Hahn, W. R.
411 Lionheart, A. Malik, R. P. Patterson, J. Stocks, A. Tizzard, N. Weiler,
412 G. K. Wolf, GREIT: A unified approach to 2D linear EIT reconstruction of
413 lung images, *Physiological Measurement* (2009). doi:10.1088/0967-3334/
414 30/6/S03.
- 415 [23] B. H. Brown, Electrical impedance tomography (EIT): A review, *Journal*
416 *of Medical Engineering and Technology* 27 (2003) 97–108. doi:10.1080/
417 0309190021000059687.
- 418 [24] N. Seifnaraghi, A. Tizzard, S. de Gelidi, D. Khodadad, S. Nordebo, A. H.
419 van Kaam, I. Frerichs, A. Waldmann, E. Sorantin, S. Tschauner, A. Demos-
420 thenous, S. Christofides, R. Bayford, Estimation of Thorax Shape for For-
421 ward Modelling in Lungs EIT, in: A. Boyle, R. Halter, E. Murphy, A. Adler
422 (Eds.), *Proceeding of the 18th International Conference on Biomedical Ap-*
423 *plications of Electrical Impedance Tomography*, Hanover (US), 2017, p. 58.
424 doi:10.5281/zenodo.557093.
- 425 [25] L. Sophocleous, I. Frerichs, M. Miedema, M. Kallio, T. Papadouri,
426 C. Karaoli, T. Becher, D. G. Tingay, A. H. Van Kaam, R. Bayford, A. D.

- 427 Waldmann, Clinical performance of a novel textile interface for neona-
428 tal chest electrical impedance tomography, *Physiological Measurement* 39
429 (2018). doi:10.1088/1361-6579/aab513.
- 430 [26] A. D. Waldmann, K. H. Wodack, A. März, A. Ukere, C. J. Trepte, S. H.
431 Böhm, D. A. Reuter, Performance of Novel Patient Interface for Electrical
432 Impedance Tomography Applications, *Journal of Medical and Biological*
433 *Engineering* 37 (2017) 561–566. doi:10.1007/s40846-017-0264-y.
- 434 [27] A. Lander, J. Newman, Paediatric anatomy, *Surgery* 31
435 (2013) 101–105. URL: [https://www.surgeryjournal.co.uk/article/](https://www.surgeryjournal.co.uk/article/S0263-9319(13)00003-3/pdf)
436 [S0263-9319\(13\)00003-3/pdf](https://www.surgeryjournal.co.uk/article/S0263-9319(13)00003-3/pdf).

Figure 1: Phantoms of simplified and identical geometry to compare the material performance: A) carbon impregnated polyurethane foam; B) mix of a silicone in two parts, CB powder and graphite powder; C) 3D printed carbon filled thermoplastic polyurethane.

A

B

Figure 2: Testing the phantoms for EIT applications: A) Ultrasound gel applied on the interface between phantoms and 32-electrode belt; B) the belt is connected to the Pioneer Set, which records the raw voltages.

Figure 3: Mean relative permittivity (dimensionless) and standard deviation bars calculated for each sample of material tested by means of the impedance analyzer.

Figure 4: EIT reconstruction of the 3D printed idealized phantom: A) Meshed geometry B) Image reconstruction.

Figure 5: 3D printed phantom of neonatal torso (A) featuring removable lungs (B).

Figure 6: Neonatal phantom of torso meshed in Comsol and imported in the EIDORS toolbox.

Figure 7: EIT analysis of the 3D printed neonatal phantom: A) Transverse view of the meshed geometry B) Image reconstruction.

Material	Conductive	Elastic*	Geometry	Adopted by
Sponge	No	Yes	Simplified	[8]
Tofu	Yes	No	Simplified	[9]
Agar	Yes	Depending on concentration	Simplified	[7, 10]
PLA	No	No	Anatomic	[11]
Mix of CB, graphite and silicone	Yes	Depending on concentration	Anatomic	[14, 18, 12, 15]
ABS and CB	Yes	No	Anatomic	[17, 18, 3]

Table 1: Literary review of materials used to generate phantoms. (*) Featuring low stiffness.